TOPO-REF-KERATOMETER RT-7000

TOPOGRAPHY, AUTO REFRACTION & KERATOMETRY

DELIGHT **IN SIGHT**

Multifunctional. Compact. Totally electronic-controlled.



- 4 systems 1 instrument
- Auto alignment + auto shot
- Colour touch screen
- Different topography maps
- Pupil + cornea ø measurement Tear stability analysis system (TSAS)



THE TOMEY TOPO-REF-KERATOMETER RT-7000



QUALITY IN DETAIL

A unique combination of Topography, Autorefraction, Keratometry and TSAS (tear stability analysis system) in one – multi diagnostic replaces four devices with all their functions. The compactness of this instrument is its strength. It is therefore a perfect space and cost saving solution for you. Highly accurate easy handling makes working with the RT-7000 professional and quick.

Colour touch screen

The 6.4 inch coloured touch screen is used as operating monimove the unit in all directions by simply touching the screen. All commands can be done via touch screen.

Auto alignment + auto shot

The handling of the RT-7000 is very easy - it does almost everything by itself. Alignment and measurement are done automati-cally. You just roughly align the system towards the patient eye and the rest is taken care of by the instrument.

SPECIFICATIONS

REFRACTIVE POWER MEASUREMENT Spherical refractive power (S) -25.00 D to +22.00 D (at VD = 12.0 mm) Measurement range

Display unit

0.01 D, 0.12 D, 0.25 D Cylindrical refractive power (C)

Measurement range Display unit

0 D to ±10.00 D (at VD = 12.0 mm) 0.01 D. 0.12 D. 0.25 D

Astigmatism axial (A) Measurement range Display unit

0° to 180° 1°

CORNEAL CURVATURE MEASUREMENT (K1, K2, AVG) Measurement range 5.00 mm to 11.00 mm / 30.68 D to 67.50 D (n=1.3375) **Display unit** 0.01 mm

CORNEAL ASTIGMATISM AND AXIS (C, A) Measurement range (C) Measurement range (A) Measurement area cornea

0 D to 10 D (n=1.3375) 0° to 180° Ø 3.0 mm (at 8.00 mm corneal curvature)

Tear stability analysis system

Tear stability analysis system for analysing the tear film stability by using the light cone system in the RT-7000. That offers you several measurement and analysing functions to detect patients with dry eye indication.

Pupil + cornea ø measurement

Once you have captured your patients eye you can set the pupil and cornea measurement bars to measure the individual diameter.

Topography indices KRI + KAI

For immediate understanding of the cornea topographic structure we have implemented the topographic indices KAI (Kerato-Asymetry Index) and KRI (Kerato-Regularity Index). These values are highligh-ted in colour (green = normal, yellow = suspect, red = abnormal) to provide you a quick information about the corneal structure behaviour.

CORNEAL SHAPE MEASUREMENT Measurement area (at 8.00 mm corneal curvature) Normal measurement mode Ø 1.0 mm to 8.0 mm Special measurement mode

PD range Minimum pupil diameter Vertex distance

MAIN UNIT

Display range

Built-in printer Output Display Chin rest

Ø 0.9 mm to 7.0 mm 9 D to 100 D 50 mm to 86 mm

Ø 2.2 mm 0 mm - 16.0 mm

Thermal printer External printer / LAN / USB 6.4" colour LCD electr. controlled

DIMENSIONS & ELECTRIC REQUIREMENTS

Dimensions WDH Weight Voltage Freauencv **Power consumption** 307 x 490 x 466 mm Annrox 20.0 kg 100 VAC to 240 VAC 50/60 Hz 120 VA to 150 VA

DIMENSIONS











Post lasik normalized man



MERCOFRAMES OPTICAL CORP 5555 Nw 74 Ave. Miami. Fl. 33166 Tel. 305-882-0120

Auto aligment

auto shot

ale@mercoframes.com